



CERTIFICATE OF ANALYSIS

Supplier Name: ALCOR Scientific
Supplier Address: 20 Thurber Blvd., Smithfield, RI, USA, 02917
Contact Phone: 800-495-5270
Email: info@alcorscientific.com

Alcor item # DSC01, DSC06
Alcor Description SEDITROL QC Material
Alcor Lot # 42
Expiration Date: 10/21/2024

Normal Control Level Tube ID: C142

Table with 3 columns: Attribute, Specification, Test Result. Rows include Mean (8-14 mm/hr, 10 mm/hr ±7), # Outside limits (0), and CV% (≤ 10% CV, 3.386%).

Abnormal Control Level Tube ID: C242

Table with 3 columns: Attribute, Specification, Test Result. Rows include Mean (55-80 mm/hr, 68 mm/hr ±28), # Outside limits (0), and CV% (≤ 10%, 0.689%).

The ALCOR Seditrol® Erythrocyte Sedimentation Rate Quality Control Material has been analyzed and certified for use with ALCOR's iSED® ESR Analyzer and miniiSED® ESR Analyzer.

Certified By: [Signature] Date: 6/15/2023