



CERTIFICATE OF ANALYSIS

Supplier Name: ALCOR® Scientific, LLC

Supplier Address: 20 Thurber Blvd., Smithfield, RI, USA, 02917

Contact Phone: 401-737-3774

Email: info@alcorscientific.com

ALCOR® Scientific item # DSC01, DSC06

Item Description SEDiTROL® ESR Controls

ALCOR® Scientific Lot # 46

Expiration Date: 22 Dec. 2025 (2025-12-22)

Normal Control Level **Tube ID: C146**

Attribute	Specification	Test Result
Mean	8 – 14 mm/hr	11 mm/hr
# Outside limits	0 (none)	0
CV%	≤ 10% CV	4.580%

Abnormal Control Level **Tube ID: C246**

Attribute	Specification	Test Result
Mean	55 – 80 mm/hr	68 mm/hr
# Outside limits	0 (none)	0
CV%	≤ 10%	0.876%

SEDiTROL® Erythrocyte Sedimentation Rate Quality Control Material has been analyzed and certified for use with ALCOR® Scientific's iSED®, iSED® ELITE, iSED PRO, and miniSED® ESR Analyzers.

Certified By: <i>J. Scott Van Alst</i>	Date: <i>6/9/2025</i>
--	-----------------------