



CERTIFICATE OF ANALYSIS

Supplier Name: ALCOR® Scientific, LLC

Supplier Address: 20 Thurber Blvd., Smithfield, RI, USA, 02917

Contact Phone: 401-737-3774

Email: info@alcorscientific.com

ALCOR® Scientific item # DSC01, DSC06

Item Description SEDiTROL® ESR Controls

ALCOR® Scientific Lot # 47

Expiration Date: 22 Sept. 2026 (2026-09-22)

Normal Control Level **Tube ID: C147**

Attribute	Specification	Test Result
Mean	8 – 14 mm/hr	13 mm/hr
# Outside limits	0 (none)	0
CV%	< 10% CV	4.753%

Abnormal Control Level **Tube ID: C247**

Attribute	Specification	Test Result
Mean	55 – 80 mm/hr	68 mm/hr
# Outside limits	0 (none)	0
CV%	< 10%	1.422%

SEDiTROL® Erythrocyte Sedimentation Rate Quality Control Material has been analyzed and certified for use with ALCOR® Scientific's iSED®, iSED® ELITE, iSED PRO, and miniiSED® ESR Analyzers.

Certified By: 	Date: <u>06/18/2025</u>
---	-------------------------