



CERTIFICATE OF ANALYSIS

Supplier Name: ALCOR® Scientific, LLC

Supplier Address: 20 Thurber Blvd., Smithfield, RI, USA, 02917

Contact Phone: 401-737-3774

Email: info@alcorscientific.com

ALCOR® Scientific item # DSC01, DSC06

Item Description SEDiTROL® ESR Controls

ALCOR® Scientific Lot # 48

Expiration Date: 2027-02-22

Normal Control Level Tube ID: C148

Attribute	Specification	Test Result
Mean	6 – 14 mm/hr	6 mm/hr
# Outside limits	0 (none)	0
CV%	≤ 10% CV	3.24%

Abnormal Control Level Tube ID: C248

Attribute	Specification	Test Result
Mean	55 – 80 mm/hr	70 mm/hr
# Outside limits	0 (none)	0
CV%	≤ 10%	0.98%

SEDiTROL® Erythrocyte Sedimentation Rate Quality Control Material has been analyzed and certified for use with ALCOR® Scientific's iSED®, iSED® ELITE, and miniiSED® ESR Analyzers.

Certified By:	J Colin Crawford	Date:	10/03/2025
---------------	------------------	-------	------------