



CERTIFICATE OF ANALYSIS

Supplier Name: ALCOR® Scientific, LLC

Supplier Address: 20 Thurber Blvd., Smithfield, RI, USA, 02917

Contact Phone: 401-737-3774

Email: info@alcorscientific.com

ALCOR® Scientific item # DSC01, DSC06

Item Description SEDiTROL® ESR Controls

ALCOR® Scientific Lot # 49

Expiration Date: 24-06-2027

Test	Specification	Result
HBsAg	Non-Reactive at Donor Level	Non-Reactive at Donor Level
HIV-1, HIV-2, Ab	Non-Reactive at Donor Level	Non-Reactive at Donor Level
HCV & HIV-1 by NAT	Non-Reactive at Donor Level	Non-Reactive at Donor Level


Normal Control Level **Tube ID:** C149

Attribute	Specification	Test Result
Mean	5 – 15 mm/hr	11 mm/hr
# Outside limits	0 (none)	0
CV%	< 10% CV	7%

Abnormal Control Level **Tube ID:** C249

Attribute	Specification	Test Result
Mean	55 – 80 mm/hr	74 mm/hr
# Outside limits	0 (none)	0
CV%	< 10%	0.96%

SEDiTROL® Erythrocyte Sedimentation Rate Quality Control Material has been analyzed and certified for use with ALCOR® Scientific's iSED®, iSED® ELITE, iSED® PRO, and miniSED® ESR Analyzers.

Certified By:	 J. Colin Crawford	Date:	02/11/2026
---------------	---	-------	------------