



**deepCLEAN™ and deepCLEAN PRO™  
Certificate of Analysis**

Supplier Name: ALCOR® Scientific LLC

Supplier Address: 20 Thurber Blvd., Smithfield, RI 02917, USA

Contact Phone: 401-737-3774

Email: [info@alcorscientific.com](mailto:info@alcorscientific.com)

ALCOR Scientific item # 112-12-022  
ALCOR Scientific Description deepCLEAN PRO  
ALCOR Scientific Job Order # JO1874  
ALCOR Scientific Lot # DC20251211  
Expiration Date: 2027-02-10  
Serial Number Range: 568 to 717

**Lot Concentration (at Dispense)**

Acceptable Range: 9% - 12% (Sodium Hypochlorite)

Test 1	Test 2	Test 3	Lot Concentration (Average of Test 1-3)
11.58 %	11.61 %	11.64 %	11.61 %

**ALCOR Scientific Tool Number:** 123-15-001

ALCOR Scientific's deepCLEAN™ has been analyzed and certified for use with ALCOR Scientific's iSED®, iSED ELITE, and miniiSED® analyzers.

ALCOR Scientific's deepCLEAN PRO™ has been analyzed and certified for use with ALCOR Scientific's iSED PRO® analyzer.

*(Concentration is expected to decrease over time until the product's expiration date due to the nature of sodium hypochlorite.)*

Certified By:	<i>Rachel Whiting</i>	Date:	12/12/2025
---------------	-----------------------	-------	------------