



CERTIFICATE OF ANALYSIS

Supplier Name: ALCOR Scientific LLC

Supplier Address: 20 Thurber Blvd., Smithfield, RI, USA, 02917

Contact Phone: 800-495-5270

Email: info@alcorscientific.com

Alcor item # 112-12-001

Alcor Description iWASH

Alcor Lot # JO1380

Expiration Date: 2026-04-22

Attribute	Specification	Test Result
Fill Volume (112-12-001)	500mL	500 mL
Fill Volume (112-12-003 & 115-12-003)	250mL	mL
Fill Volume (112-12-006)	1.8 L	L
Resistivity	16.5 – 18.2 MOhms	18.2 MOhms
TOC	≤ 20 ppb	6.85 ppb

The ALCOR Wash Solutions mentioned above have been analyzed and certified for use with ALCOR's iSED® ESR Analyzer, miniiSED® ESR Analyzer, iSED PRO ESR Analyzer, and Mizar Analyzer per the ALCOR Wash Solution Manufacturing Procedure (112-24-054) and with criteria recorded on the Milli-Q Water System Maintenance Log (112-16-097).

Certified By: 	Date: 4/25/2025
---	-----------------