



deepCLEAN™ Lot Record

Supplier Name: ALCOR® Scientific LLC

Supplier Address: 20 Thurber Blvd., Smithfield, RI 02917, USA

Contact Phone: 401-737-3774

Email: info@alcorscientific.com

ALCOR Scientific item # 112-12-020

ALCOR Scientific Description deepCLEAN

ALCOR Scientific Lot # DC20250325

Expiration Date: 2026-03-00

deepCLEAN Lot Concentration

Acceptable Range: 2.5% - 10%

Test 1	Test 2	Test 3	Lot Concentration (Average of Test 1-3)
9.23 %	9.28 %	9.28 %	9.26 %

ALCOR Scientific Tool Child Number: N/A

ALCOR Scientific's deepCLEAN has been analyzed and certified for use with ALCOR Scientific's iSED®, iSED ELITE, and miniSED® analyzers.

Certified By: 	Date: 3/25/2025
---	-----------------