



## deepCLEAN™ Lot Record

Supplier Name: ALCOR® Scientific LLC

Supplier Address: 20 Thurber Blvd., Smithfield, RI 02917, USA

Contact Phone: 401-737-3774

Email: [info@alcorscientific.com](mailto:info@alcorscientific.com)

ALCOR Scientific item # 112-12-020

ALCOR Scientific Description deepCLEAN

ALCOR Scientific Lot # DC20250404

Expiration Date: 2026-04-00


### deepCLEAN Lot Concentration

Acceptable Range: 9% - 12%

Test 1	Test 2	Test 3	Lot Concentration (Average of Test 1-3)
11.26 %	11.24 %	11.25 %	11.25 %

ALCOR Scientific Tool Child Number: N/A

ALCOR Scientific's deepCLEAN has been analyzed and certified for use with ALCOR Scientific's iSED®, iSED ELITE, and miniiSED® analyzers.

Certified By: 	Date: 4/4/2025
--	----------------