



**deepCLEAN™ and deepCLEAN PRO™
Certificate of Analysis**

Supplier Name: ALCOR® Scientific LLC

Supplier Address: 20 Thurber Blvd., Smithfield, RI 02917, USA

Contact Phone: 401-737-3774

Email: info@alcorscientific.com

ALCOR Scientific item # 112-12-020

ALCOR Scientific Description deepCLEAN

ALCOR Scientific Job Order # JO1453

ALCOR Scientific Lot # DC20250519

Expiration Date: 2026-05-00

Serial Number Range: N/A to N/A

Lot Concentration (at Dispense)

Acceptable Range: 9% - 12% (Sodium Hypochlorite)

Test 1	Test 2	Test 3	Lot Concentration (Average of Test 1-3)
9.24 %	9.25 %	9.25 %	9.2466 %

ALCOR Scientific Tool Number: 123-15-001

ALCOR Scientific's deepCLEAN™ has been analyzed and certified for use with ALCOR Scientific's iSED®, iSED ELITE, and miniiSED® analyzers.

ALCOR Scientific's deepCLEAN PRO™ has been analyzed and certified for use with ALCOR Scientific's iSED PRO® analyzer.

(Concentration is expected to decrease over time until the product's expiration date due to the nature of sodium hypochlorite.)

Certified By: <u>Bryan Feeney</u>	Date: <u>5/19/2025</u>
-----------------------------------	------------------------