



**deepCLEAN™ and deepCLEAN PRO™  
Certificate of Analysis**

Supplier Name: ALCOR® Scientific LLC

Supplier Address: 20 Thurber Blvd., Smithfield, RI 02917, USA

Contact Phone: 401-737-3774

Email: [info@alcorscientific.com](mailto:info@alcorscientific.com)

ALCOR Scientific item # 112-12-020

ALCOR Scientific Description deepCLEAN

ALCOR Scientific Job Order # JO1989

ALCOR Scientific Lot # DC20260219

Expiration Date: 2027-04-20

Serial Number Range: N/A to N/A

**Lot Concentration (at Dispense)**

Acceptable Range: 9% - 12% (Sodium Hypochlorite)

Test 1	Test 2	Test 3	Lot Concentration (Average of Test 1-3)
11.49 %	11.53 %	11.54 %	11.52 %

**ALCOR Scientific Tool Number:** 123-15-001

ALCOR Scientific's deepCLEAN™ has been analyzed and certified for use with ALCOR Scientific's iSED®, iSED ELITE, and miniiSED® analyzers.

ALCOR Scientific's deepCLEAN PRO™ has been analyzed and certified for use with ALCOR Scientific's iSED PRO® analyzer.

*(Concentration is expected to decrease over time until the product's expiration date due to the nature of sodium hypochlorite.)*

Certified By: <u>Bryan Feeney</u>	Date: <u>02/19/2026</u>
-----------------------------------	-------------------------