



### CERTIFICATE OF ANALYSIS

Supplier Name: ALCOR Scientific

Supplier Address: 20 Thurber Blvd., Smithfield, RI, USA, 02917

Contact Phone: 800-495-5270

Email: info@alcorscientific.com

Alcor item # 112-12-001, 112-12-003, 115-12-003

Alcor Description iWASH, miniiWASH, MIZARWash

Alcor Lot # 11/2024

Expiration Date: 10/2025

Attribute	Specification	Test Result
Fill Pressure	55 – 65 psi	62 psi
Fill Volume (112-12-001)	500mL	500 mL
Fill Volume (112-12-003 & 115-12-003)	250mL	250 mL
Resistivity	16.5 – 18.2 MOhms	18.2 MOhms
TOC	≤ 20 ppb	7.2 ppb

The ALCOR Wash Solutions mentioned above have been analyzed and certified for use with ALCOR's iSED® ESR Analyzer, miniiSED® ESR Analyzer and Mizar Analyzer per the ALCOR Wash Solution Manufacturing, Labeling, and Final QC Procedure (112-24-054) and criteria recorded on the Milli-Q Water System Maintenance Log (112-16-097).

Certified By:  Date: 11/22/24